


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

|  |   |  |
|--|---|--|
| <br>6950<br>Accredited to<br>ISO/IEC 17025:2005 | <b>Healthy Buildings (Ireland) Ltd</b><br><b>(trading as HBE)</b> |  |
|  | Issue No: 001      Issue date: Draft                              |  |
|  | 19 Millvale Road<br>Newry<br>Co Down<br>BT35 7NH                  | Contact: Mr A Johnson<br>Tel: +44 (0) 845 639 9673<br>Fax: +44 (0) 28 302 52269<br>E-Mail: Andrew.Johnson@hberm.com<br>Website: www.hbi.ie |
| Testing performed by the Organisation at the locations specified   |   |  |

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details  | Activity  | Location code |
|---|---|---------------|
| Liverpool Office<br>Unit 6 Century Building<br>Tower Street<br>Brunswick Business Park<br>Liverpool<br>L3 4BJ<br><br><b>Local Contact:</b> Mr A Johnson<br>Tel: +44 (0) 845 639 9673<br>E-Mail:<br>Andrew.Johnson@hberm.com | Health and Hygiene<br>Fibre Counting of asbestos<br>fibres in Air | A             |

#### Site activities performed away from the locations listed above:

| Location details   | Activity         | Location code |
|--|------------------|---------------|
| Premises including<br>domestic, commercial and<br>industrial | Health & Hygiene | B             |
| Mobile Laboratories  | Health & Hygiene | C             |



Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Healthy Buildings (Ireland) Ltd**  
**(trading as HBE)**

**Issue No: 001    Issue date: Draft**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used                              | Location Code |
|---------------------------|---|---|---------------|
| ASBESTOS FIBRES IN AIR    | <u>Health and hygiene</u>                             | HSG 248:February 2005 (Documented In-House procedure)                           |               |
|                           | Fibre counting  | By Phase contrast microscope (PCM) to HSG 248:2005 by in house procedure IP-043 | A,B,C         |
|                           | Sampling of Air for Fibre Counting                    | By in house procedure IP-043 based upon HSG 248                                 | A,B,C         |
|                           | 4 Stage Clearance Process                             | By in house procedure IM-016 based upon HSG 248                                 | B, C          |
| END                       |   |   |               |